

Conductivity/ultra-pure water measuring cells
with a twin-electrode system

Types 202922, 202923, 202924 and 202925

Glass conductivity cells, type 201080

Diaphragm tubes, type 201083

Compensation thermometer, type 201085

N cable socket, type 201090

B20.2900.01

Operating Instructions



2009-12-10/00550313



Warning

A sudden sensor malfunction could potentially result in dangerous and imprecise dosing! Suitable preventive measures must be in place to prevent this from happening.



Note

Please read these operating instructions before putting the instrument into operation. Keep the manual in a place which is accessible to all users at all times.

All the necessary settings are described in these operating instructions. If any difficulties should nevertheless arise during startup, please do not tamper with the instrument in any way. By doing so, you could endanger your rights under the instrument warranty! Please contact your supplier.



Note

Conductive conductivity cells are not authorized for use in highly adherent, oily or glutinous media - we recommend using our inductive conductivity measuring instruments here!



Note

A flat-rate charge of EUR 35 will be made if we receive instruments without a description of their fault. This fee will be added to the possible cost of repair.

1	Conductivity/ultra-pure water measuring cells with a twin-electrode system, types 202922, 202923, 202924 and 202925	2
1.1	Application	2
1.2	Principle of measurement	2
1.3	Measuring cells for laboratory and industrial use	2
1.4	Measuring ranges	2
1.5	Electrical connection	3
1.6	Installation	4
1.7	Maintenance / cleaning	5
1.8	Troubleshooting	5
1.9	Screwing the conductivity cell into the fitting	6
2	Glass conductivity cells, type 201080	8
2.1	Application	8
2.2	Technical data	9
2.3	Mounting	9
2.4	Maintenance	9
3	Diaphragm tubes, type 201083	10
3.1	Application	10
3.2	Technical data	10
3.3	Mounting	11
3.4	Maintenance	12
4	Compensation thermometers, type 201085	13
4.1	Application	13
4.2	Technical data	13
4.3	Mounting	14
5	N cable socket, type 201090	16
5.1	Application	16
5.2	Mounting	17

1 Conductivity/ultra-pure water measuring cells with a twin-electrode system, types 202922, 202923, 202924 and 202925

1.1 Application

Conductive conductivity cells are used in conjunction with suitable transmitters in industrial analysis measurement technology to determine the electrolytic conductivity of liquid media (or the resistance, in the case of ultra-pure water).

1.2 Principle of measurement

Two conductive electrodes of a defined area are immersed in the sample medium, a specific distance apart. An AC voltage of a specific measurement frequency (subject to the measuring range), is supplied to the electrodes by a separate transmitter. The conductive components (ions, salts) contained in the sample medium cause an alternating current to appear between the electrodes, which the transmitter uses to determine and display the conductivity, and convert it to a standard signal.

1.3 Measuring cells for laboratory and industrial use

Conductivity cells consists of a plastic or stainless steel flow-through, immersion or screw-in body and the embedded electrodes. Depending on the type, application and measuring range, the two electrodes are made from materials such as stainless steel, titanium, platinum or special-purpose graphite. The conductivity cells come from the manufacturer with a fixed cell constant, K [1/cm].

Typical cell constants include: $K=0.01 / 0.1 / 1.0 / 3.0$ or 10.0 . Intermediate values are possible for customized versions.

The downstream transmitter must be set to the cell constant of the measuring cell. Additional temperature sensors can be installed in the measuring cells, subject to the particular application.

1.4 Measuring ranges

The measuring range of conductive conductivity cells is physically restricted to max. 200 mS/cm.

The measuring ranges are roughly divided up according to cell constants, in the table below.



Note

The actual measuring range limits will vary, depending on the electrode material, the design and the downstream transmitter!

Cell constant K [1/cm]	Max. measuring range
0.01	up to 5 $\mu\text{S/cm}$ or 20 $\text{M}\Omega\text{cm}$
0.01	up to 10 $\mu\text{S/cm}$
0.1	up to 3000 $\mu\text{S/cm}$
1.0	up to 15 mS/cm
3.0	up to 30 mS/cm
10.0	up to 200 mS/cm

1.5 Electrical connection


The measuring cells come with a fixed cable or with a detachable plug connector, depending on the version.




Caution

The connecting cable must not be routed via the terminal blocks, but must run directly to the transmitter. Use shielded cables only, and if possible, those that are recommended / supplied by the manufacturer.

Follow the instructions in the transmitter operating manual for electrical connection!

Connection for	Instrument connector	Fixed cable	M12 connector
Outer electrode		white	1
Inner electrode	2	brown	2
Temperature compensation	1	yellow green	3
	3		4

3-wire circuit	-	-	5
Shield	-		-


1.6 Installation



Caution

Please heed the technical data for your measuring cell (see data sheets 20.2922, 20.2923, 20.2924 and 20.2925). The measuring cell must be suitable for the temperature, pressure and medium conditions specified for the system (including chemical resistance).

Do not make any mechanical modifications to the measuring cell (electrodes shortened, drilled, bent or scratched). This can result in the loss of proper functionality, as well as the rights under the instrument warranty.



Do not use a metal seal.



Note

Basically any installation position is possible. However, you must ensure that sufficient sample medium flows through and around the measuring cell (that is, the conductive measuring cell electrodes must always be completely surrounded by the medium). Structural measures must be taken to prevent flow separation or gas bubbles.

1.7 Maintenance / cleaning

The conductive conductivity cell electrodes are in direct contact with the sample medium. **Regular cleaning must therefore be performed, relative to the susceptibility of the medium to contamination!**

All suitable, common household cleaning chemicals can be used for cleaning. Abrasive cleaners have limited suitability! The measurement electrodes must not be damaged mechanically! Dilute hydrochloric acid, or cleaning in ultrasonic baths, can be helpful to prevent various accumulations, for example.

1.8 Troubleshooting

Troubleshooting must always consider all the components of the conductivity measurement chain!

The transmitter and the connecting cable must be checked, as well as the measuring cell.

Error	Possible cause	Remedy
Measurement value is too high or too low	Measuring cell is dirty	Section 1.7 "Maintenance / cleaning", page 5

No conductivity measurement (e.g. display shows "0")	Broken lead, incorrect terminal assignment. Measuring cell exposed to air (not fully immersed).	Carefully check the electrical connection again! Check the measuring cell installation location: is liquid medium present?
Error	Possible cause	Remedy
No temperature measurement (measuring cells with integrated temperature sensor)	Broken lead, incorrect electrical connection.	Carefully check the electrical connection again!
Display value unstable, fluctuating	Malfunction caused by incorrectly / insufficiently shielded connecting cable. Malfunction caused by gas bubbles.	Check the cable connection and routing. Check the installation location and position of the measuring cell and modify where necessary.



Note

The measuring cell can also be checked for short-circuits or internal contact problems. You need a continuity tester (such as the diode tester of a multimeter) to do this.

1.9 Screwing the conductivity cell into the fitting


- * Loosen the cable gland (1).
- * Run the connecting cable (3) of the conductivity cell (4) through the fitting (2).

-
- * Screw the conductivity cell (4) into the fitting (2).
Tightening torque approx. 2.5 Nm.
 - * Tighten the cable gland (1).
Tightening torque approx. 2 Nm.
-



Caution

**When removing the conductivity cell from the fitting:
First loosen the cable gland (1)!**



2 Glass conductivity cells, type 201080

2.1 Application

With type 201080/17-... glass conductivity cells, the conductivity of liquids can be determined in conjunction with a conductivity transmitter.

The parts of the measuring cell that come into contact with the sample medium are composed of glass and platinum. This ensures extensive resistance to aggressive media. The active component (the platinum electrode) can be platinized for use at higher conductivities.

The connections must be kept perfectly clean and dry, to avoid creep currents. During assembly work with coaxial cables, make sure that the black, semi-conducting layer between the braided shield and the inner insulation is removed.

All instruments and components are carefully checked before leaving the factory. Should you nevertheless have cause for complaint, please send the device back to us, free of harmful contamination. Checking returned goods is extremely complicated. It is therefore essential for you to provide more detailed information about the fault.

2.2 Technical data

Active component	Platinum
Measuring range, unplatinized	up to 1 mS/cm
Measuring range, platinized	up to 100 mS/cm
Cell constant	$k = 1 \pm 10\%$
permis. medium temperature	-10 to +160°C
Stem length	120 mm
Stem diameter	12 mm
permis. pressure	0 to 16 bar at 25°C
Connection	
- Type 201080/17-40-21-120/000, unplatinized	N plug cap
- Type 201080/17-40-21-120/000, platinized	N plug cap
- Type 201080/17-41-22-120/000, unplatinized	N screw plug cap, Pg 13.5
- Type 201080/17-40-22-120/000, platinized	N screw plug cap, Pg 13.5

2.3 Mounting

Glass conductivity cells are protected by a protective cap during delivery. This protective cap must be removed before it can be used.

Please follow the selection table for conductivity cells.

2.4 Maintenance

Dirty platinum electrodes can be cleaned by rinsing them in lye. Carefully remove stubborn deposits with a soft brush.

To minimize polarization error at high conductivities, platinized conductivity cells (recognizable from their blackened platinum surfaces) can be re-platinized. Galvanic re-platinization takes place at 20 mA, and takes four minutes in a platinization solution. Sales no. 20/00301092.

3 Diaphragm tubes, type 201083



Note

Diaphragm tubes come with three replacement diaphragms.

3.1 Application

Diaphragm tubes are used in conjunction with reference electrodes in a KCl storage vessel, as a reference system, whenever an increased electrolyte flow rate into the sample medium is required, e.g. in emulsions, varnishes and paints.

An electrolyte bridge can be formed in conjunction with a KCl storage vessel, which is connected to the diaphragm tube by a hose. Electrolyte bridges are used if the sample medium poisons the reference system, e.g. media containing sulphides and photographic chemicals.


3.2 Technical data

Material	
- Type 201080/15-87-04-22-120	PP
- Type 201080/15-88-04-22-120	PVDF
permis. medium temperature	
- Type 201080/15-87-04-22-120	-10 to +95°C
- Type 201080/15-88-04-22-120	-10 to +135°C
permis. pressure (with KCl storage vessel)	0 to 10 bar at 25°C
Diaphragm	PTFE, ø 5 mm
Stem length	120 mm
Stem diameter	12 mm
Connection	Crimp connection for PU plastic hose 8 x 6 mm ø (pressure-resistant)

3.3 Mounting

3.3.1 Screw in the diaphragm tube

The diaphragm tube can be screwed into a Pg 13.5 receiving thread. Max. tightening torque 10 Nm.




3.4 Maintenance

The flow rate can be reduced by compressing the PTFE diaphragm. The set screw is tightened to achieve this.

If a greater flow rate is subsequently required, the compressed diaphragm must be replaced with a new one. Three replacement diaphragms are included with a new diaphragm tube.

Before cleaning the diaphragm, you must check the material compatibility of the cleaning method.



4 Compensation thermometers, type 201085

4.1 Application

Compensation thermometers are used in conjunction with a relevant transmitter for temperature measurement and for automatic temperature compensation during electrochemical measurements (pH, conductivity, etc.).

4.2 Technical data

Material	Glass
permis. medium temperature	-20 to +135°C
permis. pressure	0 to 10 bar at 25°C
Stem length	120 mm
Stem diameter	12 mm
Active component	Pt100 Basic values as defined by DIN 43760, Class B
Time constant	
- T_{90}	0.8 sec
- T_{90}	4 s
Connection	
- Type 201080/16-89-1003-21-120	N plug cap
- Type 201080/16-89-1003-22-120	N screw plug cap, Pg 13.5


4.3 Mounting

4.3.1 Type 201080/16-89-1003-21-120 - plug-in


The compensation thermometer can be plugged into a 12 mm \varnothing $+0.5$ mm -0 mm receiving hole. A PVDF M12 nut, AF19, is used as a seal.

4.3.2 Type 201080/16-89-1003-22-120 - screw-in

The compensation thermometer can be screwed into a Pg 13.5 receiving thread. Max. tightening torque 10 Nm.



Type 201080/16-89-1003-21-120









Type 201080/16-89-1003-22-120

5 N cable socket, type 201090

5.1 Application

The N cable socket is intended for subsequent assembly and is **not** included in the standard scope of delivery!

The sales no. for the N cable socket is 20/00057350.



(1)	Clamping piece	(4)	Cap
(2)	Spacer sleeve	(5)	Cable guide
(3)	Set screw		

5.2 Mounting

1. Push the cap (4), cable guide (5) and spacer sleeve (2) onto the cable.
2. Strip the cable as shown in the diagram.
Warning: Remove the black, semiconducting layer!
Do not damage the cable core when stripping the cable!
3. Slide the clamping piece (1) over the braiding (shield) of the coaxial cable and apply pressure. Soft-solder the cable core with L-Sn 60 Pb Cu2 as defined by DIN 1707.
Warning: Do not use solder paste!
4. Slide the spacer sleeve (2) over the clamping piece (1), push the cable guide (5) up to the end of the spacer sleeve (2), pull the cap (4) over it and screw it firmly into the cap (4) with the set screw (3).
5. Check the complete coaxial cable for continuity and short-circuits.



JUMO GmbH & Co. KG

Street address:
Moritz-Juchheim-Straße 1
36039 Fulda, Germany
Delivery address:
Mackenrodtstraße 14
36039 Fulda, Germany
Postal address:
36035 Fulda, Germany
Phone: +49 661 6003-0
Fax: +49 661 6003-607
E-mail: mail@jumo.net
Internet: www.jumo.net

JUMO Instrument Co. Ltd.

JUMO House
Temple Bank, Riverway
Harlow - Essex CM20 2DY, UK
Phone: +44 1279 63 55 33
Fax: +44 1279 63 52 62
E-mail: sales@jumo.co.uk
Internet: www.jumo.co.uk

JUMO Process Control, Inc.

8 Technology Boulevard
Canastota, NY 13032, USA
Phone: 315-697-JUMO
1-800-554-JUMO
Fax: 315-697-5867
E-mail: info@jumo.us
Internet: www.jumo.us